



**MONITOR™
FOR CHLORINE**

(0 - 300 ppm)

QSR

Test for Free Chlorine

100 Strips

Product
Code
5148Q




© Serim Research Corporation
3506 Reedy Drive, Elkhart, IN 46514

IMPORTANT: Keep cap tight. Store at 15° - 30°C (59° - 86°F). Do not remove desiccant packet. Lot number and expiration date on bottom of bottle.

DIRECTIONS:

1. Immerse indicator pad in sample and remove *immediately*.
2. Shake strip firmly to remove excess sample.
3. Between 15 and 30 seconds after immersion, compare indicator pad to color chart. Match color at *center* of pad to the color block; disregard color around edges.

Chlorine Concentration (ppm)



MADE AND PRINTED IN THE U.S.A.
M1745148Q100 11291 R1/18